


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p>0669</p> <p>Accredited to ISO/IEC 17025:2017</p>	<h3>Novatron</h3> <p>Issue No: 023 Issue date: 21 October 2019</p>	
	<p>Novatron Scientific Ltd Novatron House 46 London Road Horsham West Sussex RH12 1AY</p>	<p>Contact: Mr S H Long Tel: +44 (0)1403 754416 Fax: +44 (0)1403 754480 E-Mail: simonhlong@btopenworld.com Website: www.novatron.co.uk</p>
<p>Calibration performed at the above address only</p>		

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks
RELATIVE HUMIDITY Relative humidity meters	11 %rh to 50 %rh 50 %rh to 90 %rh <i>for the temperature range 20 °C to 25 °C</i>	1.7 %rh 2.0 %rh	<p>Calibrations of devices with an electrical output may be undertaken.</p> <p>The accreditation covers other humidity units directly related to relative humidity, eg, equilibrium relative humidity (ERH).</p>
Saturated salt capsules	6 %rh 11 %rh 33 %rh 53 %rh 58 %rh 75 %rh 84 %rh 90 %rh 97 %rh 98 %rh	1.7 %rh 1.7 %rh 1.7 %rh 1.7 %rh 1.7 %rh 1.7 %rh 1.7 %rh 1.7 %rh 2.5 % rh 2.5 %rh	
TEMPERATURE Temperature sensors with indicators (measurements made in air)	20 °C to 25 °C	0.20 °C	Calibrations of devices with an electrical output may be undertaken.
<p>END</p>			